ShahInject BIO

INJECTOR FOR FOLDABLE INTRAOCULAR LENS

EN

WARNING

SHAHINJECT BIO is used together with IOLs from different manufacturers. Before use, ensure that the cartridge is compatible with the lens and the incision. The lens must only be loaded into the cartridge with sterile and rounded forceps SHAHINJECT BIO must not be used in combination with 3-piece lenses.

The intended Users are healthcare professionals

PRODUCT DESCRIPTION

SHAHINJECT BIO is a sterile, single-use injection system. With SHAHINJECT BIO, foldable intraocular lenses can be implanted through small incisions. The product consists of the injector and a cartridge for loading the IOL. The size must be chosen according to the lens.

The cartridge of the SHAHINJECT BIO, it is coated. This reduces the sliding resistance of the lens and thus enables easy and safe injection of the lens.

INSTRUCTION FOR USE

1 Carefully open the primary packaging and place the Contents on a sterile surface. Depending on the model, the cartridge is already mounted in the injector.

2 Completely fill the cartridge tip with viscoelastic materi

3 Apply little viscoelastic material meriat to the loading area and silicone cushion ! The viscoelastic material should be applied at least 30 seconds before injecting the IOL to ensure maximum hydration of the coating. After this time, the IOL should be injected immediately, as the lubricity decreases again!

4 Insert foldable lens with non-traumatic forceps as illustrated. Consider the instructions of the lens manufacturer.

5 Align the lens with the tweezers so that neither the haptics nor the optics are jammed when closing the cartridge

6 Close the loading chamber, press the wings together until you hear a click. Haptics and optics must not be trapped between the wings.

7 Insert the cartridge tip into the incision. Push plunger to implant lens into intended position within anterior chamber. Adjust delivery by rotating the injector. ! Do not press the cushion outside the cartridge!

Disposal

The medical de may be contaminated with potentially infectious substances of human origin after use. Dispose of the medical device and packaging after use according to the applicable guidelines of biological hazardous waste.

Notifications to manufacturer and authorities

Serious incidents occuring in connection with the device must be reported to the manufacturer and the competent authority in which the user and/or patient is established.

DE

WARNING

SHAHINJECT BIO is used together with IOLs from different manufacturers. Before use, ensure that the cartridge is compatible with the lens and the incision. The lens must only be loaded into the cartridge with sterile and rounded forceps SHAHINJECT BIO must not be used in combination with 3-piece lenses.

The intended Users are healthcare professionals

PRODUCT DESCRIPTION

SHAHINJECT BIO is a sterile, single-use injection system. With SHAHINJECT BIO, foldable intraocular lenses can be implanted through small incisions. The product consists of the injector and a cartridge for loading the IOL. The size must be chosen according to the lens.

The cartridge of the SHAHINJECT BIO, it is coated. This reduces the sliding resistance of the lens and thus enables easy and safe injection of the lens.

INSTRUCTION FOR USE

1 Carefully open the primary packaging and place the Contents on a sterile surface. Depending on the model, the cartridge is already mounted in the injector.

2 Completely fill the cartridge tip with viscoelastic materi

3 Apply little viscoelastic material meriat to the loading area and silicone cushion ! The viscoelastic material should be applied at least 30 seconds before injecting the IOL to ensure maximum hydration of the coating. After this time, the IOL should be injected immediately, as the lubricity decreases again!

4 Insert foldable lens with non-traumatic forceps as illustrated. Consider the instructions of the lens manufacturer.

5 Align the lens with the tweezers so that neither the haptics nor the optics are jammed when closing the cartridge

6 Close the loading chamber, press the wings together until you hear a click. Haptics and optics must not be trapped between the wings.

7 Insert the cartridge tip into the incision. Push plunger to implant lens into intended position within anterior chamber. Adjust delivery by rotating the injector. ! Do not press the cushion outside the cartridge!

Disposal

The medical de may be contaminated with potentially infectious substances of human origin after use. Dispose of the medical device and packaging after use according to the applicable guidelines of biological hazardous waste.

Notifications to manufacturer and authorities

Serious incidents occuring in connection with the device must be reported to the manufacturer and the competent authority in which the user and/or patient is established.

BG

WARNING

SHAHINJECT BIO is used together with IOLs from different manufacturers. Before use, ensure that the cartridge is compatible with the lens and the incision. The lens must only be loaded into the cartridge with sterile and rounded forceps SHAHINJECT BIO must not be used in combination with 3-piece lenses.

The intended Users are healthcare professionals

PRODUCT DESCRIPTION

SHAHINJECT BIO is a sterile, single-use injection system. With SHAHINJECT BIO, foldable intraocular lenses can be implanted through small incisions. The product consists of the injector and a cartridge for loading the IOL. The size must be chosen according to the lens.

The cartridge of the SHAHINJECT BIO, it is coated. This reduces the sliding resistance of the lens and thus enables easy and safe injection of the lens.

INSTRUCTION FOR USE

1 Carefully open the primary packaging and place the Contents on a sterile surface. Depending on the model, the cartridge is already mounted in the injector.

2 Completely fill the cartridge tip with viscoelastic materi

3 Apply little viscoelastic material meriat to the loading area and silicone cushion ! The viscoelastic material should be applied at least 30 seconds before injecting the IOL to ensure maximum hydration of the coating. After this time, the IOL should be injected immediately, as the lubricity decreases again!

4 Insert foldable lens with non-traumatic forceps as illustrated. Consider the instructions of the lens manufacturer.

5 Align the lens with the tweezers so that neither the haptics nor the optics are jammed when closing the cartridge

6 Close the loading chamber, press the wings together until you hear a click. Haptics and optics must not be trapped between the wings.

7 Insert the cartridge tip into the incision. Push plunger to implant lens into intended position within anterior chamber. Adjust delivery by rotating the injector. ! Do not press the cushion outside the cartridge!

Disposal

The medical de may be contaminated with potentially infectious substances of human origin after use. Dispose of the medical device and packaging after use according to the applicable guidelines of biological hazardous waste.

Notifications to manufacturer and authorities

Serious incidents occuring in connection with the device must be reported to the manufacturer and the competent authority in which the user and/or patient is established.

CN

WARNING

SHAHINJECT BIO is used together with IOLs from different manufacturers. Before use, ensure that the cartridge is compatible with the lens and the incision. The lens must only be loaded into the cartridge with sterile and rounded forceps SHAHINJECT BIO must not be used in combination with 3-piece lenses.

The intended Users are healthcare professionals

PRODUCT DESCRIPTION

SHAHINJECT BIO is a sterile, single-use injection system. With SHAHINJECT BIO, foldable intraocular lenses can be implanted through small incisions. The product consists of the injector and a cartridge for loading the IOL. The size must be chosen according to the lens.

The cartridge of the SHAHINJECT BIO, it is coated. This reduces the sliding resistance of the lens and thus enables easy and safe injection of the lens.

INSTRUCTION FOR USE

1 Carefully open the primary packaging and place the Contents on a sterile surface. Depending on the model, the cartridge is already mounted in the injector.

2 Completely fill the cartridge tip with viscoelastic materi

3 Apply little viscoelastic material meriat to the loading area and silicone cushion ! The viscoelastic material should be applied at least 30 seconds before injecting the IOL to ensure maximum hydration of the coating. After this time, the IOL should be injected immediately, as the lubricity decreases again!

4 Insert foldable lens with non-traumatic forceps as illustrated. Consider the instructions of the lens manufacturer.

5 Align the lens with the tweezers so that neither the haptics nor the optics are jammed when closing the cartridge

6 Close the loading chamber, press the wings together until you hear a click. Haptics and optics must not be trapped between the wings.

7 Insert the cartridge tip into the incision. Push plunger to implant lens into intended position within anterior chamber. Adjust delivery by rotating the injector. ! Do not press the cushion outside the cartridge!

Disposal

The medical de may be contaminated with potentially infectious substances of human origin after use. Dispose of the medical device and packaging after use according to the applicable guidelines of biological hazardous waste.

Notifications to manufacturer and authorities

Serious incidents occuring in connection with the device must be reported to the manufacturer and the competent authority in which the user and/or patient is established.

DK

WARNING

SHAHINJECT BIO is used together with IOLs from different manufacturers. Before use, ensure that the cartridge is compatible with the lens and the incision. The lens must only be loaded into the cartridge with sterile and rounded forceps SHAHINJECT BIO must not be used in combination with 3-piece lenses.

The intended Users are healthcare professionals

PRODUCT DESCRIPTION

SHAHINJECT BIO is a sterile, single-use injection system. With SHAHINJECT BIO, foldable intraocular lenses can be implanted through small incisions. The product consists of the injector and a cartridge for loading the IOL. The size must be chosen according to the lens.

The cartridge of the SHAHINJECT BIO, it is coated. This reduces the sliding resistance of the lens and thus enables easy and safe injection of the lens.

INSTRUCTION FOR USE

1 Carefully open the primary packaging and place the Contents on a sterile surface. Depending on the model, the cartridge is already mounted in the injector.

2 Completely fill the cartridge tip with viscoelastic materi

3 Apply little viscoelastic material meriat to the loading area and silicone cushion ! The viscoelastic material should be applied at least 30 seconds before injecting the IOL to ensure maximum hydration of the coating. After this time, the IOL should be injected immediately, as the lubricity decreases again!

4 Insert foldable lens with non-traumatic forceps as illustrated. Consider the instructions of the lens manufacturer.

5 Align the lens with the tweezers so that neither the haptics nor the optics are jammed when closing the cartridge

6 Close the loading chamber, press the wings together until you hear a click. Haptics and optics must not be trapped between the wings.

7 Insert the cartridge tip into the incision. Push plunger to implant lens into intended position within anterior chamber. Adjust delivery by rotating the injector. ! Do not press the cushion outside the cartridge!

Disposal

The medical de may be contaminated with potentially infectious substances of human origin after use. Dispose of the medical device and packaging after use according to the applicable guidelines of biological hazardous waste.

Notifications to manufacturer and authorities

Serious incidents occuring in connection with the device must be reported to the manufacturer and the competent authority in which the user and/or patient is established.

NL

WARNING

SHAHINJECT BIO is used together with IOLs from different manufacturers. Before use, ensure that the cartridge is compatible with the lens and the incision. The lens must only be loaded into the cartridge with sterile and rounded forceps SHAHINJECT BIO must not be used in combination with 3-piece lenses.

The intended Users are healthcare professionals

PRODUCT DESCRIPTION

SHAHINJECT BIO is a sterile, single-use injection system. With SHAHINJECT BIO, foldable intraocular lenses can be implanted through small incisions. The product consists of the injector and a cartridge for loading the IOL. The size must be chosen according to the lens.

The cartridge of the SHAHINJECT BIO, it is coated. This reduces the sliding resistance of the lens and thus enables easy and safe injection of the lens.

INSTRUCTION FOR USE

1 Carefully open the primary packaging and place the Contents on a sterile surface. Depending on the model, the cartridge is already mounted in the injector.

2 Completely fill the cartridge tip with viscoelastic materi

3 Apply little viscoelastic material meriat to the loading area and silicone cushion ! The viscoelastic material should be applied at least 30 seconds before injecting the IOL to ensure maximum hydration of the coating. After this time, the IOL should be injected immediately, as the lubricity decreases again!

4 Insert foldable lens with non-traumatic forceps as illustrated. Consider the instructions of the lens manufacturer.

5 Align the lens with the tweezers so that neither the haptics nor the optics are jammed when closing the cartridge

6 Close the loading chamber, press the wings together until you hear a click. Haptics and optics must not be trapped between the wings.

7 Insert the cartridge tip into the incision. Push plunger to implant lens into intended position within anterior chamber. Adjust delivery by rotating the injector. ! Do not press the cushion outside the cartridge!

Disposal

The medical de may be contaminated with potentially infectious substances of human origin after use. Dispose of the medical device and packaging after use according to the applicable guidelines of biological hazardous waste.

Notifications to manufacturer and authorities

Serious incidents occuring in connection with the device must be reported to the manufacturer and the competent authority in which the user and/or patient is established.

FR

WARNING

SHAHINJECT BIO is used together with IOLs from different manufacturers. Before use, ensure that the cartridge is compatible with the lens and the incision. The lens must only be loaded into the cartridge with sterile and rounded forceps SHAHINJECT BIO must not be used in combination with 3-piece lenses.

The intended Users are healthcare professionals

PRODUCT DESCRIPTION

SHAHINJECT BIO is a sterile, single-use injection system. With SHAHINJECT BIO, foldable intraocular lenses can be implanted through small incisions. The product consists of the injector and a cartridge for loading the IOL. The size must be chosen according to the lens.

The cartridge of the SHAHINJECT BIO, it is coated. This reduces the sliding resistance of the lens and thus enables easy and safe injection of the lens.

INSTRUCTION FOR USE

1 Carefully open the primary packaging and place the Contents on a sterile surface. Depending on the model, the cartridge is already mounted in the injector.

2 Completely fill the cartridge tip with viscoelastic materi

3 Apply little viscoelastic material meriat to the loading area and silicone cushion ! The viscoelastic material should be applied at least 30 seconds before injecting the IOL to ensure maximum hydration of the coating. After this time, the IOL should be injected immediately, as the lubricity decreases again!

4 Insert foldable lens with non-traumatic forceps as illustrated. Consider the instructions of the lens manufacturer.

5 Align the lens with the tweezers so that neither the haptics nor the optics are jammed when closing the cartridge

6 Close the loading chamber, press the wings together until you hear a click. Haptics and optics must not be trapped between the wings.

7 Insert the cartridge tip into the incision. Push plunger to implant lens into intended position within anterior chamber. Adjust delivery by rotating the injector. ! Do not press the cushion outside the cartridge!

Disposal

The medical de may be contaminated with potentially infectious substances of human origin after use. Dispose of the medical device and packaging after use according to the applicable guidelines of biological hazardous waste.

Notifications to manufacturer and authorities

Serious incidents occuring in connection with the device must be reported to the manufacturer and the competent authority in which the user and/or patient is established.

GR

WARNING

SHAHINJECT BIO is used together with IOLs from different manufacturers. Before use, ensure that the cartridge is compatible with the lens and the incision. The lens must only be loaded into the cartridge with sterile and rounded forceps SHAHINJECT BIO must not be used in combination with 3-piece lenses.

The intended Users are healthcare professionals

PRODUCT DESCRIPTION

SHAHINJECT BIO is a sterile, single-use injection system. With SHAHINJECT BIO, foldable intraocular lenses can be implanted through small incisions. The product consists of the injector and a cartridge for loading the IOL. The size must be chosen according to the lens.

The cartridge of the SHAHINJECT BIO, it is coated. This reduces the sliding resistance of the lens and thus enables easy and safe injection of the lens.

INSTRUCTION FOR USE

1 Carefully open the primary packaging and place the Contents on a sterile surface. Depending on the model, the cartridge is already mounted in the injector.

2 Completely fill the cartridge tip with viscoelastic materi

3 Apply little viscoelastic material meriat to the loading area and silicone cushion ! The viscoelastic material should be applied at least 30 seconds before injecting the IOL to ensure maximum hydration of the coating. After this time, the IOL should be injected immediately, as the lubricity decreases again!

4 Insert foldable lens with non-traumatic forceps as illustrated. Consider the instructions of the lens manufacturer.

5 Align the lens with the tweezers so that neither the haptics nor the optics are jammed when closing the cartridge

6 Close the loading chamber, press the wings together until you hear a click. Haptics and optics must not be trapped between the wings.

7 Insert the cartridge tip into the incision. Push plunger to implant lens into intended position within anterior chamber. Adjust delivery by rotating the injector. ! Do not press the cushion outside the cartridge!

Disposal

The medical de may be contaminated with potentially infectious substances of human origin after use. Dispose of the medical device and packaging after use according to the applicable guidelines of biological hazardous waste.

Notifications to manufacturer and authorities

Serious incidents occuring in connection with the device must be reported to the manufacturer and the competent authority in which the user and/or patient is established.

IT

WARNING

SHAHINJECT BIO is used together with IOLs from different manufacturers. Before use, ensure that the cartridge is compatible with the lens and the incision. The lens must only be loaded into the cartridge with sterile and rounded forceps SHAHINJECT BIO must not be used in combination with 3-piece lenses.

The intended Users are healthcare professionals

PRODUCT DESCRIPTION

SHAHINJECT BIO is a sterile, single-use injection system. With SHAHINJECT BIO, foldable intraocular lenses can be implanted through small incisions. The product consists of the injector and a cartridge for loading the IOL. The size must be chosen according to the lens.

The cartridge of the SHAHINJECT BIO, it is coated. This reduces the sliding resistance of the lens and thus enables easy and safe injection of the lens.

INSTRUCTION FOR USE

1 Carefully open the primary packaging and place the Contents on a sterile surface. Depending on the model, the cartridge is already mounted in the injector.

2 Completely fill the cartridge tip with viscoelastic materi

3 Apply little viscoelastic material meriat to the loading area and silicone cushion ! The viscoelastic material should be applied at least 30 seconds before injecting the IOL to ensure maximum hydration of the coating. After this time, the IOL should be injected immediately, as the lubricity decreases again!

4 Insert foldable lens with non-traumatic forceps as illustrated. Consider the instructions of the lens manufacturer.

5 Align the lens with the tweezers so that neither the haptics nor the optics are jammed when closing the cartridge

6 Close the loading chamber, press the wings together until you hear a click. Haptics and optics must not be trapped between the wings.

7 Insert the cartridge tip into the incision. Push plunger to implant lens into intended position within anterior chamber. Adjust delivery by rotating the injector. ! Do not press the cushion outside the cartridge!

Disposal

The medical de may be contaminated with potentially infectious substances of human origin after use. Dispose of the medical device and packaging after use according to the applicable guidelines of biological hazardous waste.

Notifications to manufacturer and authorities

Serious incidents occuring in connection with the device must be reported to the manufacturer and the competent authority in which the user and/or patient is established.

LV

WARNING

SHAHINJECT BIO is used together with IOLs from different manufacturers. Before use, ensure that the cartridge is compatible with the lens and the incision. The lens must only be loaded into the cartridge with sterile and rounded forceps SHAHINJECT BIO must not be used in combination with 3-piece lenses.

The intended Users are healthcare professionals

PRODUCT DESCRIPTION

SHAHINJECT BIO is a sterile, single-use injection system. With SHAHINJECT BIO, foldable intraocular lenses can be implanted through small incisions. The product consists of the injector and a cartridge for loading the IOL. The size must be chosen according to the lens.

The cartridge of the SHAHINJECT BIO, it is coated. This reduces the sliding resistance of the lens and thus enables easy and safe injection of the lens.

INSTRUCTION FOR USE

1 Carefully open the primary packaging and place the Contents on a sterile surface. Depending on the model, the cartridge is already mounted in the injector.

2 Completely fill the cartridge tip with viscoelastic materi

3 Apply little viscoelastic material meriat to the loading area and silicone cushion ! The viscoelastic material should be applied at least 30 seconds before injecting the IOL to ensure maximum hydration of the coating. After this time, the IOL should be injected immediately, as the lubricity decreases again!

4 Insert foldable lens with non-traumatic forceps as illustrated. Consider the instructions of the lens manufacturer.

5 Align the lens with the tweezers so that neither the haptics nor the optics are jammed when closing the cartridge

6 Close the loading chamber, press the wings together until you hear a click. Haptics and optics must not be trapped between the wings.

7 Insert the cartridge tip into the incision. Push plunger to implant lens into intended position within anterior chamber. Adjust delivery by rotating the injector. ! Do not press the cushion outside the cartridge!

Disposal

The medical de may be contaminated with potentially infectious substances of human origin after use. Dispose of the medical device and packaging after use according to the applicable guidelines of biological hazardous waste.

Notifications to manufacturer and authorities

Serious incidents occuring in connection with the device must be reported to the manufacturer and the competent authority in which the user and/or patient is established.

LT

WARNING

SHAHINJECT BIO is used together with IOLs from different manufacturers. Before use, ensure that the cartridge is compatible with the lens and the incision. The lens must only be loaded into the cartridge with sterile and rounded forceps SHAHINJECT BIO must not be used in combination with 3-piece lenses.

The intended Users are healthcare professionals

PRODUCT DESCRIPTION

SHAHINJECT BIO is a sterile, single-use injection system. With SHAHINJECT BIO, foldable intraocular lenses can be implanted through small incisions. The product consists of the injector and a cartridge for loading the IOL. The size must be chosen according to the lens.

The cartridge of the SHAHINJECT BIO, it is coated. This reduces the sliding resistance of the lens and thus enables easy and safe injection of the lens.

INSTRUCTION FOR USE

1 Carefully open the primary packaging and place the Contents on a sterile surface. Depending on the model, the cartridge is already mounted in the injector.

2 Completely fill the cartridge tip with viscoelastic materi

3 Apply little viscoelastic material meriat to the loading area and silicone cushion ! The viscoelastic material should be applied at least 30 seconds before injecting the IOL to ensure maximum hydration of the coating. After this time, the IOL should be injected immediately, as the lubricity decreases again!

4 Insert foldable lens with non-traumatic forceps as illustrated. Consider the instructions of the lens manufacturer.

5 Align the lens with the tweezers so that neither the haptics nor the optics are jammed when closing the cartridge

6 Close the loading chamber, press the wings together until you hear a click. Haptics and optics must not be trapped between the wings.

7 Insert the cartridge tip into the incision. Push plunger to implant lens into intended position within anterior chamber. Adjust delivery by rotating the injector. ! Do not press the cushion outside the cartridge!

Disposal

The medical de may be contaminated with potentially infectious substances of human origin after use. Dispose of the medical device and packaging after use according to the applicable guidelines of biological hazardous waste.

Notifications to manufacturer and authorities

Serious incidents occuring in connection with the device must be reported to the manufacturer and the competent authority in which the user and/or patient is established.



ShahInject BIO

INJECTOR FOR FOLDABLE INTRAOCULAR LENS

WARNING

SHAHINJECT BIO is used together with IOLs from different manufacturers. Before use, ensure that the cartridge is compatible with the lens and the incision. The lens must only be loaded into the cartridge with sterile and rounded forceps SHAHINJECT BIO must not be used in combination with 3-piece lenses.

The intended Users are healthcare professionals

PRODUCT DESCRIPTION

SHAHINJECT BIO is a sterile, single-use injection system. With SHAHINJECT BIO, foldable intraocular lenses can be implanted through small incisions. The product consists of the injector and a cartridge for loading the IOL. The size must be chosen according to the lens.

The cartridge of the SHAHINJECT BIO, it is coated. This reduces the sliding resistance of the lens and thus enables easy and safe injection of the lens.

INSTRUCTION FOR USE

1 Carefully open the primary packaging and place the Contents on a sterile surface. Depending on the model, the cartridge is already mounted in the injector.

2 Completely fill the cartridge tip with viscoelastic materi

3 Apply little viscoelastic material meriel to the loading area and silicone cushion. ! The viscoelastic material should be applied at least 30 seconds before injecting the IOL to ensure maximum hydration of the coating. After this time, the IOL should be injected immediately, as the lubricity decreases again!

4 Insert foldable lens with non-traumatic forceps as illustrated. Consider the instructions of the lens manufacturer.

5 Align the lens with the tweezers so that neither the haptics nor the optics are jammed when closing the cartridge

6 Close the loading chamber, press the wings together until you hear a click. Haptics and optics must not be trapped between the wings.

7 Insert the cartridge tip into the incision. Push plunger to implant lens into intended position within anterior chamber. Adjust delivery by rotating the injector. ! Do not press the cushion outside the cartridge!

Disposal

The medical de may be contaminated with potentially infectious substances of human origin after use. Dispose of the medical device and packaging after use according to the applicable guidelines of biological hazardous waste.

Notifications to manufacturer and authorities

Serious incidents occuring in connection with the device must be reported to the manufacturer and the competent authority in which the user and/or patient is established.

LOT Lot number · Expiry date · Keep dry · Keep away from sunlight · Date of Manufacture · European representative · Do not use if package is damaged · Do not reuse · Do not resterilize · Consult instructions for use · Single sterile barrier system · Unique Device Identifier · Manufacturer · For Professionals Only (FDA symbol) · Sterile by Ethylene oxide (EO) · Medical Device · Caution · Conformité Européenne

PT

WARNING

SHAHINJECT BIO is used together with IOLs from different manufacturers. Before use, ensure that the cartridge is compatible with the lens and the incision. The lens must only be loaded into the cartridge with sterile and rounded forceps SHAHINJECT BIO must not be used in combination with 3-piece lenses.

The intended Users are healthcare professionals

PRODUCT DESCRIPTION

SHAHINJECT BIO is a sterile, single-use injection system. With SHAHINJECT BIO, foldable intraocular lenses can be implanted through small incisions. The product consists of the injector and a cartridge for loading the IOL. The size must be chosen according to the lens.

The cartridge of the SHAHINJECT BIO, it is coated. This reduces the sliding resistance of the lens and thus enables easy and safe injection of the lens.

INSTRUCTION FOR USE

1 Carefully open the primary packaging and place the Contents on a sterile surface. Depending on the model, the cartridge is already mounted in the injector.

2 Completely fill the cartridge tip with viscoelastic materi

3 Apply little viscoelastic material meriel to the loading area and silicone cushion. ! The viscoelastic material should be applied at least 30 seconds before injecting the IOL to ensure maximum hydration of the coating. After this time, the IOL should be injected immediately, as the lubricity decreases again!

4 Insert foldable lens with non-traumatic forceps as illustrated. Consider the instructions of the lens manufacturer.

5 Align the lens with the tweezers so that neither the haptics nor the optics are jammed when closing the cartridge

6 Close the loading chamber, press the wings together until you hear a click. Haptics and optics must not be trapped between the wings.

7 Insert the cartridge tip into the incision. Push plunger to implant lens into intended position within anterior chamber. Adjust delivery by rotating the injector. ! Do not press the cushion outside the cartridge!

Disposal

The medical de may be contaminated with potentially infectious substances of human origin after use. Dispose of the medical device and packaging after use according to the applicable guidelines of biological hazardous waste.

Notifications to manufacturer and authorities

Serious incidents occuring in connection with the device must be reported to the manufacturer and the competent authority in which the user and/or patient is established.

LOT Lot number · Expiry date · Keep dry · Keep away from sunlight · Date of Manufacture · European representative · Do not use if package is damaged · Do not reuse · Do not resterilize · Consult instructions for use · Single sterile barrier system · Unique Device Identifier · Manufacturer · For Professionals Only (FDA symbol) · Sterile by Ethylene oxide (EO) · Medical Device · Caution · Conformité Européenne

RO

WARNING

SHAHINJECT BIO is used together with IOLs from different manufacturers. Before use, ensure that the cartridge is compatible with the lens and the incision. The lens must only be loaded into the cartridge with sterile and rounded forceps SHAHINJECT BIO must not be used in combination with 3-piece lenses.

The intended Users are healthcare professionals

PRODUCT DESCRIPTION

SHAHINJECT BIO is a sterile, single-use injection system. With SHAHINJECT BIO, foldable intraocular lenses can be implanted through small incisions. The product consists of the injector and a cartridge for loading the IOL. The size must be chosen according to the lens.

The cartridge of the SHAHINJECT BIO, it is coated. This reduces the sliding resistance of the lens and thus enables easy and safe injection of the lens.

INSTRUCTION FOR USE

1 Carefully open the primary packaging and place the Contents on a sterile surface. Depending on the model, the cartridge is already mounted in the injector.

2 Completely fill the cartridge tip with viscoelastic materi

3 Apply little viscoelastic material meriel to the loading area and silicone cushion. ! The viscoelastic material should be applied at least 30 seconds before injecting the IOL to ensure maximum hydration of the coating. After this time, the IOL should be injected immediately, as the lubricity decreases again!

4 Insert foldable lens with non-traumatic forceps as illustrated. Consider the instructions of the lens manufacturer.

5 Align the lens with the tweezers so that neither the haptics nor the optics are jammed when closing the cartridge

6 Close the loading chamber, press the wings together until you hear a click. Haptics and optics must not be trapped between the wings.

7 Insert the cartridge tip into the incision. Push plunger to implant lens into intended position within anterior chamber. Adjust delivery by rotating the injector. ! Do not press the cushion outside the cartridge!

Disposal

The medical de may be contaminated with potentially infectious substances of human origin after use. Dispose of the medical device and packaging after use according to the applicable guidelines of biological hazardous waste.

Notifications to manufacturer and authorities

Serious incidents occuring in connection with the device must be reported to the manufacturer and the competent authority in which the user and/or patient is established.

LOT Lot number · Expiry date · Keep dry · Keep away from sunlight · Date of Manufacture · European representative · Do not use if package is damaged · Do not reuse · Do not resterilize · Consult instructions for use · Single sterile barrier system · Unique Device Identifier · Manufacturer · For Professionals Only (FDA symbol) · Sterile by Ethylene oxide (EO) · Medical Device · Caution · Conformité Européenne

RU

WARNING

SHAHINJECT BIO is used together with IOLs from different manufacturers. Before use, ensure that the cartridge is compatible with the lens and the incision. The lens must only be loaded into the cartridge with sterile and rounded forceps SHAHINJECT BIO must not be used in combination with 3-piece lenses.

The intended Users are healthcare professionals

PRODUCT DESCRIPTION

SHAHINJECT BIO is a sterile, single-use injection system. With SHAHINJECT BIO, foldable intraocular lenses can be implanted through small incisions. The product consists of the injector and a cartridge for loading the IOL. The size must be chosen according to the lens.

The cartridge of the SHAHINJECT BIO, it is coated. This reduces the sliding resistance of the lens and thus enables easy and safe injection of the lens.

INSTRUCTION FOR USE

1 Carefully open the primary packaging and place the Contents on a sterile surface. Depending on the model, the cartridge is already mounted in the injector.

2 Completely fill the cartridge tip with viscoelastic materi

3 Apply little viscoelastic material meriel to the loading area and silicone cushion. ! The viscoelastic material should be applied at least 30 seconds before injecting the IOL to ensure maximum hydration of the coating. After this time, the IOL should be injected immediately, as the lubricity decreases again!

4 Insert foldable lens with non-traumatic forceps as illustrated. Consider the instructions of the lens manufacturer.

5 Align the lens with the tweezers so that neither the haptics nor the optics are jammed when closing the cartridge

6 Close the loading chamber, press the wings together until you hear a click. Haptics and optics must not be trapped between the wings.

7 Insert the cartridge tip into the incision. Push plunger to implant lens into intended position within anterior chamber. Adjust delivery by rotating the injector. ! Do not press the cushion outside the cartridge!

Disposal

The medical de may be contaminated with potentially infectious substances of human origin after use. Dispose of the medical device and packaging after use according to the applicable guidelines of biological hazardous waste.

Notifications to manufacturer and authorities

Serious incidents occuring in connection with the device must be reported to the manufacturer and the competent authority in which the user and/or patient is established.

LOT Lot number · Expiry date · Keep dry · Keep away from sunlight · Date of Manufacture · European representative · Do not use if package is damaged · Do not reuse · Do not resterilize · Consult instructions for use · Single sterile barrier system · Unique Device Identifier · Manufacturer · For Professionals Only (FDA symbol) · Sterile by Ethylene oxide (EO) · Medical Device · Caution · Conformité Européenne

SK

WARNING

SHAHINJECT BIO is used together with IOLs from different manufacturers. Before use, ensure that the cartridge is compatible with the lens and the incision. The lens must only be loaded into the cartridge with sterile and rounded forceps SHAHINJECT BIO must not be used in combination with 3-piece lenses.

The intended Users are healthcare professionals

PRODUCT DESCRIPTION

SHAHINJECT BIO is a sterile, single-use injection system. With SHAHINJECT BIO, foldable intraocular lenses can be implanted through small incisions. The product consists of the injector and a cartridge for loading the IOL. The size must be chosen according to the lens.

The cartridge of the SHAHINJECT BIO, it is coated. This reduces the sliding resistance of the lens and thus enables easy and safe injection of the lens.

INSTRUCTION FOR USE

1 Carefully open the primary packaging and place the Contents on a sterile surface. Depending on the model, the cartridge is already mounted in the injector.

2 Completely fill the cartridge tip with viscoelastic materi

3 Apply little viscoelastic material meriel to the loading area and silicone cushion. ! The viscoelastic material should be applied at least 30 seconds before injecting the IOL to ensure maximum hydration of the coating. After this time, the IOL should be injected immediately, as the lubricity decreases again!

4 Insert foldable lens with non-traumatic forceps as illustrated. Consider the instructions of the lens manufacturer.

5 Align the lens with the tweezers so that neither the haptics nor the optics are jammed when closing the cartridge

6 Close the loading chamber, press the wings together until you hear a click. Haptics and optics must not be trapped between the wings.

7 Insert the cartridge tip into the incision. Push plunger to implant lens into intended position within anterior chamber. Adjust delivery by rotating the injector. ! Do not press the cushion outside the cartridge!

Disposal

The medical de may be contaminated with potentially infectious substances of human origin after use. Dispose of the medical device and packaging after use according to the applicable guidelines of biological hazardous waste.

Notifications to manufacturer and authorities

Serious incidents occuring in connection with the device must be reported to the manufacturer and the competent authority in which the user and/or patient is established.

LOT Lot number · Expiry date · Keep dry · Keep away from sunlight · Date of Manufacture · European representative · Do not use if package is damaged · Do not reuse · Do not resterilize · Consult instructions for use · Single sterile barrier system · Unique Device Identifier · Manufacturer · For Professionals Only (FDA symbol) · Sterile by Ethylene oxide (EO) · Medical Device · Caution · Conformité Européenne

SL

WARNING

SHAHINJECT BIO is used together with IOLs from different manufacturers. Before use, ensure that the cartridge is compatible with the lens and the incision. The lens must only be loaded into the cartridge with sterile and rounded forceps SHAHINJECT BIO must not be used in combination with 3-piece lenses.

The intended Users are healthcare professionals

PRODUCT DESCRIPTION

SHAHINJECT BIO is a sterile, single-use injection system. With SHAHINJECT BIO, foldable intraocular lenses can be implanted through small incisions. The product consists of the injector and a cartridge for loading the IOL. The size must be chosen according to the lens.

The cartridge of the SHAHINJECT BIO, it is coated. This reduces the sliding resistance of the lens and thus enables easy and safe injection of the lens.

INSTRUCTION FOR USE

1 Carefully open the primary packaging and place the Contents on a sterile surface. Depending on the model, the cartridge is already mounted in the injector.

2 Completely fill the cartridge tip with viscoelastic materi

3 Apply little viscoelastic material meriel to the loading area and silicone cushion. ! The viscoelastic material should be applied at least 30 seconds before injecting the IOL to ensure maximum hydration of the coating. After this time, the IOL should be injected immediately, as the lubricity decreases again!

4 Insert foldable lens with non-traumatic forceps as illustrated. Consider the instructions of the lens manufacturer.

5 Align the lens with the tweezers so that neither the haptics nor the optics are jammed when closing the cartridge

6 Close the loading chamber, press the wings together until you hear a click. Haptics and optics must not be trapped between the wings.

7 Insert the cartridge tip into the incision. Push plunger to implant lens into intended position within anterior chamber. Adjust delivery by rotating the injector. ! Do not press the cushion outside the cartridge!

Disposal

The medical de may be contaminated with potentially infectious substances of human origin after use. Dispose of the medical device and packaging after use according to the applicable guidelines of biological hazardous waste.

Notifications to manufacturer and authorities

Serious incidents occuring in connection with the device must be reported to the manufacturer and the competent authority in which the user and/or patient is established.

LOT Lot number · Expiry date · Keep dry · Keep away from sunlight · Date of Manufacture · European representative · Do not use if package is damaged · Do not reuse · Do not resterilize · Consult instructions for use · Single sterile barrier system · Unique Device Identifier · Manufacturer · For Professionals Only (FDA symbol) · Sterile by Ethylene oxide (EO) · Medical Device · Caution · Conformité Européenne

ES

WARNING

SHAHINJECT BIO is used together with IOLs from different manufacturers. Before use, ensure that the cartridge is compatible with the lens and the incision. The lens must only be loaded into the cartridge with sterile and rounded forceps SHAHINJECT BIO must not be used in combination with 3-piece lenses.

The intended Users are healthcare professionals

PRODUCT DESCRIPTION

SHAHINJECT BIO is a sterile, single-use injection system. With SHAHINJECT BIO, foldable intraocular lenses can be implanted through small incisions. The product consists of the injector and a cartridge for loading the IOL. The size must be chosen according to the lens.

The cartridge of the SHAHINJECT BIO, it is coated. This reduces the sliding resistance of the lens and thus enables easy and safe injection of the lens.

INSTRUCTION FOR USE

1 Carefully open the primary packaging and place the Contents on a sterile surface. Depending on the model, the cartridge is already mounted in the injector.

2 Completely fill the cartridge tip with viscoelastic materi

3 Apply little viscoelastic material meriel to the loading area and silicone cushion. ! The viscoelastic material should be applied at least 30 seconds before injecting the IOL to ensure maximum hydration of the coating. After this time, the IOL should be injected immediately, as the lubricity decreases again!

4 Insert foldable lens with non-traumatic forceps as illustrated. Consider the instructions of the lens manufacturer.

5 Align the lens with the tweezers so that neither the haptics nor the optics are jammed when closing the cartridge

6 Close the loading chamber, press the wings together until you hear a click. Haptics and optics must not be trapped between the wings.

7 Insert the cartridge tip into the incision. Push plunger to implant lens into intended position within anterior chamber. Adjust delivery by rotating the injector. ! Do not press the cushion outside the cartridge!

Disposal

The medical de may be contaminated with potentially infectious substances of human origin after use. Dispose of the medical device and packaging after use according to the applicable guidelines of biological hazardous waste.

Notifications to manufacturer and authorities

Serious incidents occuring in connection with the device must be reported to the manufacturer and the competent authority in which the user and/or patient is established.

LOT Lot number · Expiry date · Keep dry · Keep away from sunlight · Date of Manufacture · European representative · Do not use if package is damaged · Do not reuse · Do not resterilize · Consult instructions for use · Single sterile barrier system · Unique Device Identifier · Manufacturer · For Professionals Only (FDA symbol) · Sterile by Ethylene oxide (EO) · Medical Device · Caution · Conformité Européenne

SV

WARNING

SHAHINJECT BIO is used together with IOLs from different manufacturers. Before use, ensure that the cartridge is compatible with the lens and the incision. The lens must only be loaded into the cartridge with sterile and rounded forceps SHAHINJECT BIO must not be used in combination with 3-piece lenses.

The intended Users are healthcare professionals

PRODUCT DESCRIPTION

SHAHINJECT BIO is a sterile, single-use injection system. With SHAHINJECT BIO, foldable intraocular lenses can be implanted through small incisions. The product consists of the injector and a cartridge for loading the IOL. The size must be chosen according to the lens.

The cartridge of the SHAHINJECT BIO, it is coated. This reduces the sliding resistance of the lens and thus enables easy and safe injection of the lens.

INSTRUCTION FOR USE

1 Carefully open the primary packaging and place the Contents on a sterile surface. Depending on the model, the cartridge is already mounted in the injector.

2 Completely fill the cartridge tip with viscoelastic materi

3 Apply little viscoelastic material meriel to the loading area and silicone cushion. ! The viscoelastic material should be applied at least 30 seconds before injecting the IOL to ensure maximum hydration of the coating. After this time, the IOL should be injected immediately, as the lubricity decreases again!

4 Insert foldable lens with non-traumatic forceps as illustrated. Consider the instructions of the lens manufacturer.

5 Align the lens with the tweezers so that neither the haptics nor the optics are jammed when closing the cartridge

6 Close the loading chamber, press the wings together until you hear a click. Haptics and optics must not be trapped between the wings.

7 Insert the cartridge tip into the incision. Push plunger to implant lens into intended position within anterior chamber. Adjust delivery by rotating the injector. ! Do not press the cushion outside the cartridge!

Disposal

The medical de may be contaminated with potentially infectious substances of human origin after use. Dispose of the medical device and packaging after use according to the applicable guidelines of biological hazardous waste.

Notifications to manufacturer and authorities

Serious incidents occuring in connection with the device must be reported to the manufacturer and the competent authority in which the user and/or patient is established.

LOT Lot number · Expiry date · Keep dry · Keep away from sunlight · Date of Manufacture · European representative · Do not use if package is damaged · Do not reuse · Do not resterilize · Consult instructions for use · Single sterile barrier system · Unique Device Identifier · Manufacturer · For Professionals Only (FDA symbol) · Sterile by Ethylene oxide (EO) · Medical Device · Caution · Conformité Européenne

TR

WARNING

SHAHINJECT BIO is used together with IOLs from different manufacturers. Before use, ensure that the cartridge is compatible with the lens and the incision. The lens must only be loaded into the cartridge with sterile and rounded forceps SHAHINJECT BIO must not be used in combination with 3-piece lenses.

The intended Users are healthcare professionals

PRODUCT DESCRIPTION

SHAHINJECT BIO is a sterile, single-use injection system. With SHAHINJECT BIO, foldable intraocular lenses can be implanted through small incisions. The product consists of the injector and a cartridge for loading the IOL. The size must be chosen according to the lens.

The cartridge of the SHAHINJECT BIO, it is coated. This reduces the sliding resistance of the lens and thus enables easy and safe injection of the lens.

INSTRUCTION FOR USE

1 Carefully open the primary packaging and place the Contents on a sterile surface. Depending on the model, the cartridge is already mounted in the injector.

2 Completely fill the cartridge tip with viscoelastic materi

3 Apply little viscoelastic material meriel to the loading area and silicone cushion. ! The viscoelastic material should be applied at least 30 seconds before injecting the IOL to ensure maximum hydration of the coating. After this time, the IOL should be injected immediately, as the lubricity decreases again!

4 Insert foldable lens with non-traumatic forceps as illustrated. Consider the instructions of the lens manufacturer.

5 Align the lens with the tweezers so that neither the haptics nor the optics are jammed when closing the cartridge

6 Close the loading chamber, press the wings together until you hear a click. Haptics and optics must not be trapped between the wings.

7 Insert the cartridge tip into the incision. Push plunger to implant lens into intended position within anterior chamber. Adjust delivery by rotating the injector. ! Do not press the cushion outside the cartridge!

Disposal

The medical de may be contaminated with potentially infectious substances of human origin after use. Dispose of the medical device and packaging after use according to the applicable guidelines of biological hazardous waste.

Notifications to manufacturer and authorities

Serious incidents occuring in connection with the device must be reported to the manufacturer and the competent authority in which the user and/or patient is established.

LOT Lot number · Expiry date · Keep dry · Keep away from sunlight · Date of Manufacture · European representative · Do not use if package is damaged · Do not reuse · Do not resterilize · Consult instructions for use · Single sterile barrier system · Unique Device Identifier · Manufacturer · For Professionals Only (FDA symbol) · Sterile by Ethylene oxide (EO) · Medical Device · Caution · Conformité Européenne

JP

WARNING

SHAHINJECT BIO is used together with IOLs from different manufacturers. Before use, ensure that the cartridge is compatible with the lens and the incision. The lens must only be loaded into the cartridge with sterile and rounded forceps SHAHINJECT BIO must not be used in combination with 3-piece lenses.

The intended Users are healthcare professionals

PRODUCT DESCRIPTION

SHAHINJECT BIO is a sterile, single-use injection system. With SHAHINJECT BIO, foldable intraocular lenses can be implanted through small incisions. The product consists of the injector and a cartridge for loading the IOL. The size must be chosen according to the lens.

The cartridge of the SHAHINJECT BIO, it is coated. This reduces the sliding resistance of the lens and thus enables easy and safe injection of the lens.

INSTRUCTION FOR USE

1 Carefully open the primary packaging and place the Contents on a sterile surface. Depending on the model, the cartridge is already mounted in the injector.

2 Completely fill the cartridge tip with viscoelastic materi

3 Apply little viscoelastic material meriel to the loading area and silicone cushion. ! The viscoelastic material should be applied at least 30 seconds before injecting the IOL to ensure maximum hydration of the coating. After this time, the IOL should be injected immediately, as the lubricity decreases again!

4 Insert foldable lens with non-traumatic forceps as illustrated. Consider the instructions of the lens manufacturer.

5 Align the lens with the tweezers so that neither the haptics nor the optics are jammed when closing the cartridge

6 Close the loading chamber, press the wings together until you hear a click. Haptics and optics must not be trapped between the wings.

7 Insert the cartridge tip into the incision. Push plunger to implant lens into intended position within anterior chamber. Adjust delivery by rotating the injector. ! Do not press the cushion outside the cartridge!

Disposal

The medical de may be contaminated with potentially infectious substances of human origin after use. Dispose of the medical device and packaging after use according to the applicable guidelines of biological hazardous waste.

Notifications to manufacturer and authorities

Serious incidents occuring in connection with the device must be reported to the manufacturer and the competent authority in which the user and/or patient is established.

LOT Lot number · Expiry date · Keep dry · Keep away from sunlight · Date of Manufacture · European representative · Do not use if package is damaged · Do not reuse · Do not resterilize · Consult instructions for use · Single sterile barrier system · Unique Device Identifier · Manufacturer · For Professionals Only (FDA symbol) · Sterile by Ethylene oxide (EO) · Medical Device · Caution · Conformité Européenne



Manufactured by

Shah Eye Care Pvt Ltd
the wheels of change...

Plot No: 8 and 9, Karur Industrial Area,
PB Road, Davangere 577006,
Karnataka, India
Phone: +91- 8192-252151, +91 - 8192-252161
Email: sales@shaheyecare.com
Web : www.shaheyecare.com
Mfg. Lic. No:

EC REP

Advinia Ltd. Tower Business Centre, 2nd Fl.,
Tower Street, Swatar BKR 4013 Malta